



花岗石检测平台

Granite Inspection Plate

Specifications 规格	Plateness Tolerance (μm) 平面度公差				After-Packing VOL (mm) 包装后体积	Weight (kg)重量
	Grade 000	Grade 00	Grade 0	Grade 1		
160x100x50	0.8	1.5	3	6	270x210x150	8
160x160x50	0.8	1.5	3	6	270x270x150	9
250x160x50	0.9	1.8	3.5	7	370x270x150	11
250x250x50	0.9	1.8	3.5	7	370x370x150	13
300x200x50	0.9	1.8	3.6	7	410x270x180	15
300x300x70	1	1.9	3.8	7.5	410x370x200	27
400x250x70	1	2	4	8	510x470x200	30
400x400x70	1.2	2.3	4.5	9	528x488x210	46
630x400x100	1.3	2.5	5	10	740x490x270	94
630x630x100	1.3	2.5	5	10	750x730x270	155
800x500x100	1.4	2.7	5.3	11	920x600x280	155
900x600x100	1.5	2.9	5.7	11.5	1000x660x200	200
1000x630x150	1.5	3	6	12	1120x740x330	335
1000x750x150	1.6	3.2	6.3	12.5	1250x950x330	398
1000x1000x150	1.8	3.5	7	14	1250x1230x330	532
1600x1000x200	2	4	8	16	1870x1230x370	1080
1600x1600x200	2.4	4.8	9.5	19	1870x1830x370	1730
2000x1000x200	2.4	4.8	9.5	19	2270x1230x400	1345
2000x1600x250	2.5	5	10	20	2310x1830x460	2622
2500x1600x300	2.9	5.8	11.5	23	2810x1830x510	3872
3000x2000x400	3.3	6.5	13	27	3350x2230x670	7651

华安精密花岗石平板精度参照各先进国家标准, 平面度公差依: JB/T7975-1999 标准

000 级=1×(1+d/1000) μm

00 级=2×(1+/1000) μm

0 级=4×(1+d/1000) μm

1 级=8×(1+d/1000) μm



深圳市华安精密仪器配件有限公司
香港华安精密机械有限公司



(d 為對角線 mm)

(測量溫度一般在 $20 \pm 2^{\circ} \text{C}$)

花崗石平板精度表見上頁。



备注(Notes): 规格 Specifications (mm);精度等级 Precision grade 000 级;平面度 Grade 000 flatness(μm)00 级;平面度 Grade 00 Flatness(μm) 0 级;平面度 Grade 0 Flatness(μm)1 级;平面度 Grade 1 Flatness(μm) ;Net 净重 kg

TEL 电话: 0755-28795989.; 84676729; 28790149。 Fax 传真: 0755-28759846. Mail(邮箱):

baohui6@163.com ; hasales@163.com QQ: 603868280; 2231955408; 1715609993;936231399.

Website:公司网站: www.ha199.com Telephone of the Dongguan sales company 东莞销售公司联

系电话: 0769-81622875

Fax 传真: 0769-81622875

QQ: 1715609993